

## Certificate Of Analysis

**Whatman™**

### Product Information

Product Number: 1820-055

Product Name: GF/A 5.5CM 100/PK

Lot Number/Serial Number: 17258236

Date of Manufacture: 28-Jan-2021

### Quality Control

Test	Range	Results
Grammage,gsm	47-59	50
Thickness, um@53kPa	225-295	245
Porosity, S/300ml/1.0sq.in	8-18	14
Air retention	≥99.95	99.99

### Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

### Electronic signature

This document has electronically produced and is valid without a signature.

Version AB



[www.gelifesciences.com](http://www.gelifesciences.com)

**Global Life Sciences Solutions Operations UK Ltd.**

Amersham Place Little Chalfont  
Buckinghamshire HP7 9NA UK